

Certificate of analysis no: 1X/XXX

Emission date	Month X th , 201X	Internal order	XXX
Customer name	XXX	Customer order	XXX
Reference	XXX		
Description	XXX		
Length	XX	Quantity	X
Hose Batch	XXX	Silicone Batch	XXX
Manufacturing date	XX th , Month 201X	Expiration date (According ISO 2230 "Guidelines for storage")	10 years

Properties	Unit	Limits	Results
Hardness	Shore A	50 +/- 5	XX
Tensile strength	N/mm2	>5	XX
Elongation	%	>200	XX
Tear strength B	kN/m	>20	XX
Density	g/cm ₃	1,13 +/- 0,02	XX

Certified information

Process	Quality management system certified ISO 9001:2015, ISO 14001:2015 and EMAS. Produced in ISO7 classified clean room according to ISO14644.
---------	--

	FDA Standard 21 CFR 177.2600
Product	BfR Standard part XV
	USP In Vivo <88> Class VI at 121°C
	ISO 10993-5 & USP In vitro <87>
	ISO 10993-4, 6 & 10
	European Pharmacopoeia 3.1.9

QHSE approval

It is the user's responsibility to ensure the suitability and safety of the VENAIR products for all intended uses. All the tests must be conducted in accordance with applicable regulatory requirements to determine the safety and effectiveness for use of the hoses in any particular application. User assumes all other risk, if any, including the risk of injury, loss or damage, direct or consequential, arising out of the use, misuse, or inability to use, this product. Vena, Venair and the Venair logo are trademarks of Venair Ibérica SA.